

Toulouse, April 1st 2020

TEST REPORT N° 20-1441

STUDY 20 - 2616

Standard NF T 72-281 (November 2014)
Determination of virucidal activity for aerial surface disinfection processes
Human Coronavirus

Medical area

Clean condition / Obligatory conditions

Promotor

MICROBECLEAN
1 bis rue des Longrais
35520 LA-CHAPELLE-DES-FOUGERETZ

Test Laboratory

FONDEREPHAR
Faculté des Sciences Pharmaceutiques
35 Chemin des Maraîchers
31062 TOULOUSE cedex 9

Dr. Laila HADDIOUI
Assay Manager



Dr Jocelyne BACARIA
Quality Manager



1. Test Laboratory

Fondation pour le Développement de la recherche en Pharmacie (FONDEREPHAR)
Faculté des Sciences Pharmaceutiques, 35 chemin des Maraîchers 31062 Toulouse cedex 9,
France

2. Identification of the aerial disinfection system

Product : **MICROBECARE™ 70-2**
Batch : L091902 A
Expiry date : Not precised
Date of receipt : 09/01/2020
Internal code : **20-2616-1**

Active(s) substance(s): Octadecylaminodimethyltrimethoxysilylpropylammonium
CAS 27668-52-8 Isopropanol CAS 67-63-0

Device : **VP200ES (VICTORY)**
Serial number : 24475
Date of receipt : 09/01/2020
Internal code : **20-2616-2**

Device pressure (manual use) : 8 bar - nozzle 40μ
Distance device/carrier : 80cm
Waiting time : 60 minutes
Room volume : 30 to 150 m³

Promotor : **MICROBECLEAN**

Storage conditions : Ambient temperature, darkness
Period of testing : March 2020

3- Experimental conditions

3-1 Virus/Receiving cells

Virus

Name: Human Coronavirus 229E
Origin : ATCC
ATCC reference: VR-740
Batch number supplier: 58505270
Internal number Batch: SS-1-110214 (passage N°1)

Receiving cells

Origin : ATCC
Designation : VERO cells
ATCC reference: CCL-81
Batch number ATCC: 3372621
Internal number Batch: WCB-041113 (passage N°33)

3-2 Carriers

The selected tests surfaces are stainless steel discs, flats, corresponding to the requirements of paragraph 5.2.3.1 of the standard. The suppliers are MERCIER CLAUSSE.

3-3 Conditions of disinfection system use

- Room :

Relative humidity: start of trial 51% - end of trial 48% (requirements 40 - 80%).
Temperature: start of trial 21.7°C - end of trial 21.3°C (requirements 18 - 22°C).
Test room volume: 32m³

- Carriers :

The carriers were placed in a vertical position, towards the device.

- Amount of disinfectant diffusion

According to the VICTORY VP200ES device features indicated by the manufacturer, the 40μ nozzle corresponds to a 3,4 oz/minute product diffusion.

According to the Promotor specifications: diffusion in a single pass, with a distance device/carrier of: 80cm - waiting time 60 minutes

3-4 Interfering substance and culture media

-Interfering substance: BSA fraction V at 0,3g/l (Batch N°287)
-Culture media: EMEM 2% SVF (Batch N°2495)

4- Validations Protocol

4-1 Control of sensitivity of cells to virus

- Add one volume of solution S or PBS + one volume of cellular suspension at 2.10⁵ cells/ml for one hour in water bath at 36°C±1°C
- The cells are centrifuged and resuspended in culture media
- The virus is diluted from 1/4 to 1/4 on a 96-well microplate (15 dilutions)
- Add 100 μl of cell suspension treated (Solution S) or not treated (PBS control) to each well of the microplate
- Incubate for 72 hours

The difference of titre reduction between cells treated by the solution S and cells treated by PBS shall be $< 1 \lg$.

4-2 Control of efficiency for suppression of disinfectant activity

- Add 1 volume of BSA + 1 volume of virus suspension + 1 volume of solution S or distilled water
- Leave the mixture in the ice bath for 60 min at room temperature

5- Titration method

- Titrate the virus (method titration on cell in suspension) by following steps:
- Serial dilutions (1/4) are realized with culture medium in the glass tube
- Transfer 0,1 ml of each dilution into eight wells of a microplate plaque
- The last row of eight wells will receive 0,1 ml of culture medium (control untreated cells)
- Add 0,1 ml of cell suspension at $2 \cdot 10^5$ cell/ml.
- Incubate for 48 or 72 hours at $36^\circ \text{C} \pm 1^\circ \text{C}$ under $5\% \text{CO}_2 \pm 2\%$.
- The viral cytopathic effect is read by using an inverted microscope

The estimated of infectious unite is determined by method KARBBER-SPAERMAN calculating the negative logarithm of 50% endpoint ($\lg \text{DICT}_{50}$) by the following formula:

$\lg \text{DICT}_{50} = \text{negative logarithm of the highest concentration of virus} - [(\text{Sum of \% affected to each dilution}/100 - 0.5) \times (\lg \text{dilution})]$

6- Results

Virus suspension title assay: lgDICT50 8.25

No cytotoxicity was observed on the carrier without treatment which has been pretreated with the aerial disinfection system according to treatment (assay conditions: 1/50).

	Degree of cytopathogenic effect (log)	Logarithmic reduction
Sensitivity of cells to virus		
- With treatment (S1)		
Carrier 1	8.40	Difference <1 lg.
Carrier 2	8.55	
Average	8.48	
- Without traitement (S2)		
Carrier 1	7.72	
Efficiency for suppression of disinfectant activity		
- With treatment (D1)		
Carrier1	8.40	Difference <0,5 lg.
Carrier 2	8.25	
Average	8.33	
- Without traitement (D2)		
Carrier 1	7.95	
Test control		
Carrier1	6.23	
Carrier 2	6.23	
Average	6.23	
Assay		
Support 1	2.10	4.13
Support 2	2.10	
Support 3	2.10	
Average	2.10	

7- Conclusion

According to the conditions of test for the standard NF T 72-281 (November 2014), the couple device/product: VICTORY VP200ES N° serial 24475/MICROBECARE™ 70-2 Lot L091902A (Exp. not precised), for a use in medical area under clean condition, leds to a virucidal activity against Human Coronavirus 229E (log reduction ≥ 4), after a treatment with the 40 μ m nozzle - distance 80 cm - waiting time 60 minutes.